

MULTI-ENZYMATIC CLEANSER

Multi-Enzymatic Cleanser effectively cleans away blood, proteins, fat, carbohydrates, and synthetic liquids from surgical instruments, flexible and rigid scopes, and other general health care equipment. Its non-abrasive, non-foaming formula is easy to use and will not harm the equipment being cleaned. Multi-Enzymatic Cleanser is effective in both manual and automated reprocessing systems, as well as ultrasonic units and washer sterilizer/decontaminators.

DIRECTIONS FOR USE:

- Manual Cleaning:** Use ½ to 1-ounce of Multi-Enzymatic Cleanser per one gallon water for general purpose cleaning. Additional amounts may be required for hard-to-remove matter. The enzymes in the Multi-Enzymatic Cleanser have an optimum efficiency when used at a temperature of 95°F (35°C). It is recommended that these products be used between 68-122°F (20-50°C). The enzymes in the Multi-Enzymatic Cleanser may become ineffective in water temperatures exceeding this level. Multi-Enzymatic Cleanser will begin breaking down organic matter immediately. A minimum soak time of 2 minutes is recommended. Hard-to-remove matter may possibly require additional soak time.
- Automatic Cleaning Units:** Follow recommendations under Manual Cleaning except that endoscope reprocessors may require less Multi Enzymatic Cleanser. For endoscope reprocessors, ultrasonic units and washer sterilizers/decontaminators, process instruments according to manufacturer's instructions.
- Thoroughly rinse Multi-Enzymatic Cleanser from instruments making sure to aspirate rinse water through all channels, box locks, lumens, etc.
- Once instruments have dried, sterilization or disinfection may follow according to appropriate procedures.
- Only fresh, unused Multi-Enzymatic Cleanser solution should be used. Used solution may be disposed of into sewage disposal system flushing with large quantities of water and in accordance with federal, state or local regulations.
- DO NOT add other chemicals such as bleaches, peroxides or other detergents which could destroy the effectiveness of the Multi-Enzymatic Cleanser.

STORAGE AND DISPOSAL

It is recommended that Multi-Enzymatic Cleanser be stored between 32°F (0°C) and 100°F (38°C). KEEP FROM FREEZING. Thoroughly rinse container before disposing or offering for recycling.



WARNING

Causes serious eye irritation

PRECAUTIONARY STATEMENTS- PREVENTION

Wash face, hands and any exposed skin thoroughly after handling.
Wear eye/face protection

PRECAUTIONARY STATEMENTS- RESPONSE

IF IN EYES: Rinse cautiously with water for several minutes.
Remove contact lenses, if present and easy to do.
Continue rinsing and get medical attention if irritation occurs

FIRST AID MEASURES

Eye Contact	Rinse immediately with plenty of water, also under the eyelids, for at least 15 minutes. Get medical attention if needed.
Skin Contact	Wash hands thoroughly after handling.
Inhalation	Remove victim to fresh air and keep at rest in a position comfortable for breathing. If experiencing respiratory symptoms: Call Poison Control or doctor/physician.
Ingestion	Dilute by giving a large amount of water. Allow vomiting to occur, then get medical attention.

KEEP OUT OF REACH OF CHILDREN